

NEBRASKA

Good Life. Great Mission.

DEPT. OF HEALTH AND HUMAN SERVICES



Pete Ricketts, Governor

June 15, 2017

To: All VFC Providers that use the LogTag and VFC 400 dataloggers

Yesterday the LogTag Company sent an email out with information regarding an update to the LogTag data loggers. This email was sent by the company that manufactures the blue data loggers. If this software is used, the red VFC 400 dataloggers will not work. The red dataloggers that have been provided by the Nebraska Immunization Program are from Control Solutions and need to have the software developed by the company. They will not function with the software from the LogTag Company. If you updated to the version sent in the email, you will need to reinstall the Control Solution version.

The Control Solution version is located at: www.vfcdataloggers.com. Choose software downloads. Select VFC 400 software (VTMC). When the software is opened you will see at the top left corner – Control Solutions VTMC 2.7 Build 8.

You can disable the version update in your LogTag software by choosing: edit-options-update, unchecking the box, press OK.

We apologize for any confusion this may have caused from the LogTag Company, as these are sent without connection to the Nebraska Immunization Program. If there is an update to the Control Solutions software, the Nebraska Immunization Program will send out that information as it is available.

If you have any questions, please contact your state nurse representative.



Helping People Live Better Lives